





# **PRODUCT CONFORMITY CERTIFICATE**

This is to certify that the

## GA5000 Portable Gas Analyser

Manufactured by:

### QED Environmental Systems Inc.

2355 Bishop Circle West Dexter, MI 48130 USA

has been assessed by CSA Group and for the conditions stated on this certificate complies with:

MCERTS - Performance Standard for Handheld Emission Monitoring Systems (HEMs), Version 4, September 2018

Certification ranges:

$CH_4$	0 to 70% vol
CO <sub>2</sub>	0 to 60% vol
O <sub>2</sub>	0 to 21% vol
CO	0 to 2000ppm
H₂S	0 to 200ppm
Flow	0 to 20 L/h

Project No.: Certificate No: Initial Certification: This Certificate issued: Renewal Date: 80237634 CSA MC130238/06 22 October 2013 12 February 2025 11 February 2030

Andrew Young Environmental Team Manager

MCERTS is operated on behalf of the Environment Agency by

CSA Group Testing UK Ltd Unit 6, Hawarden Industrial Park Hawarden, Deeside, CH5 3US Tel: +44 (0)1244 670 900

The MCERTS certificate consists of this document in its entirety. For conditions of use, please consider all the information within. This certificate may only be reproduced in its entirety and without change To authenticate the validity of this certificate please visit www.csagroupuk.org/mcerts

Form 1335 Issue 12 Page 1 of 7







#### **Certificate Contents**

Approved Site Application	2
Basis of Certification	
Product Certified	2
Certified Performance	3
Description	7
General Notes	

#### **Approved Site Application**

Any potential user should make sure, in consultation with the manufacturer, that the monitoring system is suitable for the intended application. For further information on stack emissions monitoring refer to the Environment Agency's guidance available at <u>www.mcerts.net</u>

#### **Basis of Certification**

This certification is based on the following test report(s) and on CSA Group's assessment and ongoing surveillance of the product and the manufacturing process:

Sira Report 16A27549 dated 25/09/2013

#### **Product Certified**

The GA5000 portable gas analyser measuring system consists of the following parts:

- Dual wavelength infra-red cell with reference channel for the detection of CO<sub>2</sub> and CH<sub>4</sub>.
- Internal electrochemical cell for the detection of O<sub>2</sub>, CO and H<sub>2</sub>S.
- · Pressure sensor for the detection of gas flow.

This certificate applies to all instruments fitted with software/firmware version V1.10.6 (serial number G500020 onwards).







#### **Certified Performance**

The instrument was evaluated for use under the following conditions:

Ambient Temperature Range: +5°C to +40°C

Unless otherwise stated the evaluation was carried out on the following certification ranges:  $CH_4$  0 to 70% vol,  $CO_2$  0 to 60% vol,  $O_2$  0 to 21% vol, CO 0 to 2000ppm,  $H_2S$  0 to 200ppm, Flow 0 to 20 L/h.

Results are expressed as % of the certification range unless otherwise stated (Note 1)

Test	Resu		ssed as %		Other results	MCERTS specification
	<0.5	<1	<2	<5		specification
Warm up time					61s	Clause 5.2.2
					015	To be reported
Response time						Clause 5.2.3
$CH_4$					10s	<60s
CO <sub>2</sub>					10s	<60s
O <sub>2</sub>					23s	<200s
CO					51s	<60s
H₂S					28s	<90s
Flow					4s	<60s
Repeatability at zero point						Clause 5.2.4
$CH_4$	0.03					<±5%
CO <sub>2</sub>	0.02					<±5%
O <sub>2</sub> (note 1)	0.02					<±0.4% vol
CO	0.03					<±5%
H <sub>2</sub> S	0.2					<±5%
Flow	-					-
Repeatability at span point						Clause 5.2.5
$CH_4$	0.04					<±5%
CO <sub>2</sub>	0.17					<±5%
O <sub>2</sub> (note 1)	0.1					<±0.4% vol
CO	0.27					<±5%
$H_2S$		1.28				<±5%
Flow	0.1					<±5%

Certificate No: This certificate issued: MC130238/06 12 February 2025







Test	Resul		sed as %		Other results	MCERTS specification
	<0.5	<1	<2	<5		•
Lack-of-fit (linearity)						Clause 5.2.6
$CH_4$			1.27			<±5%
CO <sub>2</sub>		-0.72				<±5%
O <sub>2</sub> (note 1)	0.09					<±0.4% vol
со		-0.63				<±5%
H₂S			1.45			<±5%
Flow	0.47					<±5%
Influence of ambient temperature zero point						Clause 5.2.7
+5°C to +40°C						
CH <sub>4</sub>	0.11					<±10%
CO <sub>2</sub>	0.17					<±10%
O <sub>2</sub> (note 1)	-0.12					<±1.5% vol
со	0.18					<±10%
$H_2S$				-2.36		<±10%
Flow				-4.14		<±5%
Influence of ambient temperature span point						Clause 5.2.7
+5°C to +40°C						
CH <sub>4</sub>			-1.24			<±10%
CO <sub>2</sub>		0.55				<±10%
O <sub>2</sub> (note 1)	0.3					<±1.5% vol
СО					7.07	<±10%
H₂S					5.38	<±10%
Flow				-3.36		<±5%

MC130238/06 12 February 2025







Test	Results expressed as % of the certification range				Other results	MCERTS specification
	<0.5	<1	<2	<5		
Cross-sensitivity at zero						Clause 5.2.8
CH <sub>4</sub>	0.0					<±5%
CO <sub>2</sub>	0.0					<±5%
O <sub>2</sub> (note 1)	-0.14					<±0.8% vol
СО	0.0					<±5%
H <sub>2</sub> S			1.63			<±5%
Flow	-					-
Cross-sensitivity at span						Clause 5.2.8
CH <sub>4</sub>				4.25		<±5%
CO <sub>2</sub>				-4.07		<±5%
O <sub>2</sub> (note 1)		-0.53				<±0.8% vol
со		0.86				<±5%
H <sub>2</sub> S			1.93			<±5%
Flow	-					-
Zero drift – 1 hour						Clause 5.2.9
CH <sub>4</sub>	0.06					<±3%
CO <sub>2</sub>	0.03					<±3%
O <sub>2</sub> (note 1)	0.0					<±0.3% vol
со	0.04					<±3%
H <sub>2</sub> S	0.47					<±3%
Flow	-					-

MC130238/06 12 February 2025







Test		Results expressed as % of the certification range			Other results	MCERTS specification
	<0.5	<1	<2	<5		
Span drift – 1 hour						
CH <sub>4</sub>	0.03					<±3%
CO <sub>2</sub>	0.39					<±3%
O <sub>2</sub> (note 1)	0.15					<±0.3% vol
со	-0.19					<±3%
H <sub>2</sub> S			-1.78			<±3%
Flow	-					-

Note 1 The performance specifications for oxygen are expressed as a percentage of volume concentration of oxygen

Certificate No: This certificate issued: MC130238/06 12 February 2025







#### Description

The GA5000 portable gas analyser operates on the principle of infra-red absorption for the measurement of  $CO_2$  and  $CH_4$ . The radiation from a broad band IR source is passed through filters to select only the wavelengths that will be absorbed by  $CO_2$  and  $CH_4$ . A gas sample is pumped into the measurement cell where the IR radiation is passed through the gas onto a detector. The IR source is pulsed to improve signal to noise ratio. A separate reference beam is used to compensate for any instrumental drift. The measurement is compensated for temperature and pressure changes.

 $O_2$ , CO and  $H_2S$  are measured by electrochemical cells.

Flow is calculated using a pressure sensor to measure the differential pressure across a restrictor.

All data is digitised and manipulated, displayed and stored digitally. Readings can be stored and downloaded for further analysis.

The GA5000 Portable Gas Analyser can measure other gases and parameters that are not included in the certification.

#### **General Notes**

- 1. This certificate is based upon the equipment tested. The Manufacturer is responsible for ensuring that on-going production complies with the standard(s) and performance criteria defined in this certificate. The manufacturer is required to maintain an approved quality management system controlling the manufacture of the certified product. Both the product and the quality management system shall be subject to regular surveillance according to 'Regulations Applicable to the Holders of CSA Group Testing UK Ltd Certificates'.
- 2. The design of the product certified is defined in the CSA Group design schedule for certificate No. CSA MC130238/06.
- 3. If a certified product is found not to comply, CSA Group should be notified immediately at the address shown on this certificate.
- 4. The certification marks that can be applied to the product or used in publicity material are defined in 'Regulations Applicable to the Holders of CSA Group Testing UK Ltd Certificates'.
- 5. This document remains the property of CSA Group and shall be returned when requested by CSA Group.