

PRODUCT CONFORMITY CERTIFICATE

This is to certify that the

BIOGAS 5000 Portable Gas Analyser

Manufactured by:

QED Environmental Systems Inc.

2355 Bishop Circle West
Dexter, MI 48130
USA

has been assessed by CSA Group
and for the conditions stated on this certificate complies with:

MCERTS - Performance Standard for Handheld Emission Monitoring Systems (HEMs), Version 4, September 2018

Certification ranges:

CH ₄	0 to 70% vol
CO ₂	0 to 60% vol
O ₂	0 to 21% vol
CO	0 to 2000ppm
H ₂ S	0 to 200ppm

Project No.: 80237634
Certificate No: CSA MC130240/06
Initial Certification: 22 October 2013
This Certificate issued: 12 February 2025
Renewal Date: 11 February 2030



Andrew Young
Environmental Team Manager

MCERTS is operated on behalf of the Environment Agency by

CSA Group Testing UK Ltd

Unit 6, Hawarden Industrial Park
Hawarden, Deeside, CH5 3US
Tel: +44 (0)1244 670 900



0011

The MCERTS certificate consists of this document in its entirety.

For conditions of use, please consider all the information within.

This certificate may only be reproduced in its entirety and without change

To authenticate the validity of this certificate please visit www.csagroupuk.org/mcerts

Certificate Contents

Approved Site Application.....	2
Basis of Certification	2
Product Certified.....	2
Certified Performance	3
Description.....	6
General Notes	6

Approved Site Application

Any potential user should make sure, in consultation with the manufacturer, that the monitoring system is suitable for the intended application. For further information on stack emissions monitoring refer to the Environment Agency's guidance available at www.mcerts.net

Basis of Certification

This certification is based on the following test report(s) and on CSA Group's assessment and ongoing surveillance of the product and the manufacturing process:

Report 16A27549 dated 25/09/2013

Product Certified

The BIOGAS 5000 portable gas analyser measuring system consists of the following parts:

- Dual wavelength infra-red cell with reference channel for the detection of CO₂ and CH₄.
- Internal electrochemical cell for the detection of O₂, CO and H₂S.

This certificate applies to all instruments fitted with software/firmware version V1.10.6 (serial number G500020 onwards).

Certificate No: CSA MC130240/06
This certificate issued: 12 February 2025

*This certificate may only be reproduced in its entirety and without change
To authenticate the validity of this certificate please visit www.csagroupuk.org/mcerts*

Certified Performance

The instrument was evaluated for use under the following conditions:

Ambient Temperature Range: +5°C to +40°C

Unless otherwise stated the evaluation was carried out on the following certification ranges: CH₄ 0 to 70% vol, CO₂ 0 to 60% vol, O₂ 0 to 21% vol, CO 0 to 2000ppm, H₂S 0 to 5000ppm

Results are expressed as error % of certification range, unless otherwise stated.

Test	Results expressed as % of the certification range				Other results	MCERTS specification
	<0.5	<1	<2	<5		
Warm up time					59s	Clause 5.2.2 To be reported
Response time						Clause 5.2.3
CH ₄					10s	<60s
CO ₂					10s	<60s
O ₂					23s	<200s
CO					51s	<60s
H ₂ S					42s	<90s
Repeatability at zero point						Clause 5.2.4
CH ₄	0.03					<±5%
CO ₂	0.02					<±5%
O ₂ (note 1)	0.02					<±0.4% vol
CO	0.03					<±5%
H ₂ S	0.3					<±5%
Repeatability at span point						Clause 5.2.5
CH ₄	0.04					<±5%
CO ₂	0.17					<±5%
O ₂ (note 1)	0.1					<±0.4% vol
CO	0.27					<±5%
H ₂ S		0.64				<±5%

Certificate No: CSA MC130240/06
This certificate issued: 12 February 2025

*This certificate may only be reproduced in its entirety and without change
To authenticate the validity of this certificate please visit www.csagroupuk.org/mcerts*

Test	Results expressed as % of the certification range				Other results	MCERTS specification
	<0.5	<1	<2	<5		
Lack-of-fit (linearity)						Clause 5.2.6
CH ₄			1.27			<±5%
CO ₂		-0.72				<±5%
O ₂ (note 1)	0.09					<±0.4% vol
CO		-0.63				<±5%
H ₂ S		-0.94				<±5%
Influence of ambient temperature zero point +5°C to +40°C						Clause 5.2.7
CH ₄	0.11					<±10%
CO ₂	0.17					<±10%
O ₂ (note 1)	-0.12					<±1.5% vol
CO	0.18					<±10%
H ₂ S				-2.36		<±10%
Influence of ambient temperature span point +5°C to +40°C						Clause 5.2.7
CH ₄			-1.24			<±10%
CO ₂		0.55				<±10%
O ₂ (note 1)	0.3					<±1.5% vol
CO					7.07	<±10%
H ₂ S					5.38	<±10%

Certificate No: CSA MC130240/06
 This certificate issued: 12 February 2025

*This certificate may only be reproduced in its entirety and without change
 To authenticate the validity of this certificate please visit www.csagroupuk.org/mcerts*

Test	Results expressed as % of the certification range				Other results	MCERTS specification
	<0.5	<1	<2	<5		
Cross-sensitivity at zero						Clause 5.2.8
CH ₄	0.0					<±5%
CO ₂	0.0					<±5%
O ₂ (note 1)	-0.14					<±0.8% vol
CO	0.0					<±5%
H ₂ S			1.63			<±5%
Cross-sensitivity at span						Clause 5.2.8
CH ₄				4.25		<±5%
CO ₂				-4.07		<±5%
O ₂ (note 1)		-0.53				<±0.8% vol
CO		0.86				<±5%
H ₂ S			1.93			<±5%
Zero drift – 1 hour						Clause 5.2.9
CH ₄	0.06					<±3%
CO ₂	0.03					<±3%
O ₂ (note 1)	0.0					<±0.3% vol
CO	0.04					<±3%
H ₂ S	0.47					<±3%
Span drift – 1 hour						
CH ₄	0.03					<±3%
CO ₂	0.39					<±3%
O ₂ (note 1)	0.15					<±0.3% vol
CO	-0.19					<±3%
H ₂ S			-1.78			<±3%

Note 1 The performance specifications for oxygen are expressed as a percentage of volume concentration of oxygen

Certificate No: CSA MC130240/06
 This certificate issued: 12 February 2025

*This certificate may only be reproduced in its entirety and without change
 To authenticate the validity of this certificate please visit www.csagroupuk.org/mcerts*

Description

The BIOGAS 5000 portable gas analyser operates on the principle of infra-red absorption for the measurement of CO₂ and CH₄. The radiation from a broad band IR source is passed through filters to select only the wavelengths that will be absorbed by CO₂ and CH₄. A gas sample is pumped into the measurement cell where the IR radiation is passed through the gas onto a detector. The IR source is pulsed to improve signal to noise ratio. A separate reference beam is used to compensate for any instrumental drift. The measurement is compensated for temperature and pressure changes.

O₂, CO and H₂S are measured by electrochemical cells.

All data is digitised and manipulated, displayed and stored digitally. Readings can be stored and downloaded for further analysis.

The BIOGAS 5000 Portable Gas Analyser can measure other gases and parameters that are not included in the certification.

General Notes

1. This certificate is based upon the equipment tested. The Manufacturer is responsible for ensuring that on-going production complies with the standard(s) and performance criteria defined in this certificate. The manufacturer is required to maintain an approved quality management system controlling the manufacture of the certified product. Both the product and the quality management system shall be subject to regular surveillance according to 'Regulations Applicable to the Holders of CSA Group Testing UK Ltd Certificates'.
2. The design of the product certified is defined in the CSA Group design schedule for certificate No. CSA MC130240/06.
3. If a certified product is found not to comply, CSA Group should be notified immediately at the address shown on this certificate.
4. The certification marks that can be applied to the product or used in publicity material are defined in 'Regulations Applicable to the Holders of CSA Group Testing UK Ltd Certificates'.
5. This document remains the property of CSA Group and shall be returned when requested by CSA Group.

Certificate No: CSA MC130240/06
This certificate issued: 12 February 2025

*This certificate may only be reproduced in its entirety and without change
To authenticate the validity of this certificate please visit www.csagroupuk.org/mcerts*